



















concentration of duplex stainless steel modified carbon paste electrode show maximum current sensitivity. High selectivity, easy preparation and high sensitivity of the voltammetric response made DMCPCE a very useful sensor in determining DA, AA and UA. DMCPCE can be used as sensor in medical field for the diagnosis of DA, AA and UA deficiency diseases. Almost all the electrode reactions at the above bioactive compounds are adsorption controlled reactions.

## ACKNOWLEDGMENT

Financial support for this work from the Council of Scientific & Industrial Research (CSIR), India (Grant No. 22/561/11/EMR II Dated 11.04.2011) is gratefully acknowledged.

## REFERENCES

- [1] B.N. Chandrashekar, B.E. Kumara Swamy, M. Pandurangachar, T.V. Sathisha, B.S. Sherigara, "Electropolymerisation of L-arginine at carbon paste electrode and its application to the detection of dopamine, ascorbic and uric acid", *Colloid Surface B*. Vol. 88, pp. 413–418, 2011.
- [2] S. Marceglia, G. Foffani, A. Bianchi, G. Baselli, F. Tamma, M. Egidi, A. Priori, "Dopamine dependent non-linear correlation between subthalamic rhythms in Parkinson's disease", *J. Physiol*. Vol. 571, pp. 579–591, 2006.
- [3] N. Kempainen, P. Marjamaki, M. Roytta, J.O. Rinne, "Different pattern of reduction of striatal dopamine reuptake sites in Alzheimer's disease and ageing", *J. Neural Transm*. Vol. 108, pp. 827–836, 2001.
- [4] L. Schwieler, G. Engberg, S. Erhardt, "Clozapine modulates midbrain dopamine neuron firing via interaction with the NMDA receptor complex", *Synapse*. Vol. 52, pp. 114–122, 2004.
- [5] B.N. Chandrashekar, B.E. Kumara Swamy, M. Pandurangachar, S. Sharath Shankar, Ongera Gilbert, J.G. Manjunatha, B.S. Sherigara, "Electrochemical Oxidation of Dopamine at Polyethylene glycol Modified Carbon Paste Electrode: A Cyclic Voltammetric Study", *Int. J. Electrochem. Sci*. Vol. 5, pp. 578–592, 2010.
- [6] J.G. Manjunatha, B.E. Kumara Swamy, G.P. Mamatha, Ongera Gilbert, B.N. Chandrashekar, B.S. Sherigara, "Electrochemical studies of dopamine and epinephrine at a poly (tannic acid) modified carbon paste electrode: a cyclic voltammetric study", *Int. J. Electrochem. Sci*. Vol. 5, pp. 1236–1245, 2010.
- [7] K.C. Lin, P.C. Yeh, S.M. Chen, "Electrochemical determination of ascorbic acid using poly (xanthurenic acid) and multi-walled carbon nanotubes", *Int. J. Electrochem. Sci*. Vol. 7, pp. 12752–12763, 2012.
- [8] G. Hu, Y. Guo, Q. Xue, S. Shao, "A highly selective amperometric sensor for ascorbic acid based on mesopore-rich active carbon-modified pyrolytic graphite electrode", *Electrochim. Acta*. Vol. 55, pp. 2799–2804, 2010.
- [9] The Riordan IVC Protocol for Adjunctive Cancer Care: Intravenous ascorbate as a chemotherapeutic and biological response modifying agent, Riordan Clinic Research Institute. February, pp. 1–21, 2013.
- [10] A.B. Florou, M.I. Prodromidis, M.I. Karayannis, S.M. Tzouvara-Karayanni, "Flow electrochemical determination of ascorbic acid in real samples using a glassy carbon electrode modified with a cellulose acetate film bearing 2,6-dichlorophenolindophenol", *Anal. Chimica. Acta*. Vol. 409, pp. 113–121, 2000.
- [11] M.S.U. Ali, N. H. Alvi, Z. Ibupoto, O. Nur, M. Willander, B. Danielsson, "Selective potentiometric determination of uric acid with uricase immobilized on ZnO nanowires", *Sensor Actuat B Chem*. Vol. 152, pp. 241–247, 2011.
- [12] R. Shashanka, B.E. Kumara Swamy, S. Reddy, D. Chaira, "Synthesis of Silver Nanoparticles and their Applications", *Anal. Bioanal. Electrochem*. Vol. 5, pp. 455–466, 2013.
- [13] R. Shashanka, D. Chaira, "Phase transformation and microstructure study of nano structured austenitic and ferritic stainless steel powders prepared by planetary milling", *Powder Technol*. Vol. 259, pp. 125–136, 2014.
- [14] R. Shashanka, D. Chaira, "Development of nano-structured duplex and ferritic stainless steels by pulverisette planetary milling followed by pressureless sintering", *Mater Charact*. Vol. 99, pp. 220–229, 2015.
- [15] R. Shashanka, D. Chaira, "Optimization of milling parameters for the synthesis of nano-structured duplex and ferritic stainless steel powders by high energy planetary milling", *Powder Technol*. Vol. 278, pp. 35–45, 2015.
- [16] B.D. Cullity, S.R. Stock, *Elements of X-Ray Diffraction*, Pearson, 2003. (Paperback, ISBN-13: 9780131788183).
- [17] S. Brunauer, P.H. Emmett, E.J. Teller, "Adsorption of gases in multimolecular layers", *J. Am. Chem. Soc*. Vol. 60, pp. 309–319, 1938.
- [18] T. Khalil, F. Abou El-Nour, B. El-Gammal, A.R. Boccaccini, "Determination of surface area and porosity of sol-gel derived ceramic powders in the system TiO<sub>2</sub>-SiO<sub>2</sub>-Al<sub>2</sub>O<sub>3</sub>", *Powder Technol*. Vol. 114, pp. 106–111, 2001.
- [19] S.J. Gregg, K.S. Sing, "Adsorption, Surface Area and Porosity", Academic Press, London, 1982.
- [20] J.G. Manjunatha, B.E. Kumara Swamy, G.P. Mamatha, Ongera Gilbert, M.T. Shreenivas, B.S. Sherigara, "Electrocatalytic Response of Dopamine at Mannitol and Triton X-100 Modified Carbon Paste Electrode: A Cyclic Voltammetric Study", *Int. J. Electrochem. Sci*. Vol. 4, pp. 1706–1718, 2009.
- [21] J.G. Manjunatha, B.E. Kumara Swamy, R. Deepa, V. Krishna, G.P. Mamatha, Umesh Chandra, S. Sharath Shankar, B.S. Sherigara, "Electrochemical studies of Dopamine at Eperisone and Cetyl Trimethyl Ammonium Bromide Surfactant modified Carbon paste electrode: A Cyclic Voltammetric Study", *Int. J. Electrochem. Sci*. Vol. 4, pp. 662–671, 2009.
- [22] R. Shashanka, D. Chaira, B.E. Kumara Swamy, "Electrocatalytic Response of Duplex and Yittria Dispersed Duplex Stainless Steel Modified Carbon Paste Electrode in Detecting Folic Acid Using Cyclic Voltammetry", *Int. J. Electrochem. Sci*. Vol. 10, pp. 5586 – 5598, 2015.
- [23] S. Mahshid, C. Li, S.S. Mahshid, M. Askari, A. Dolati, L. Yang, S. Luoa, Q. Cai, "Sensitive determination of dopamine in the presence of uric acid and ascorbic acid using TiO<sub>2</sub> nanotubes modified with Pd, Pt and Au nanoparticles", *Analyst*. Vol. 136, pp. 2322–2329, 2011.
- [24] B.J. Venton, M. Wightman, "Psychoanalytical electrochemistry: dopamine and behavior", *Anal. Chem*. Vol. 75, pp. 414A–421A, 2003.
- [25] F. Martinello, E.L. Da Silva, "Ascorbic acid interference in the measurement of serum biochemical parameters: in vivo and in vitro studies", *Clin. Biochem*. Vol. 39, pp. 396–403, 2006.
- [26] P.S. Cahill, Q.D. Walker, J.M. Finnegan, G.E. Mickelson, E.R. Travis, R.M. Wightman, "Microelectrodes for the measurement of catecholamines in biological systems", *Anal. Chem*. Vol. 68, pp. 3180–3186, 1996.
- [27] O. Gilbert, B.E.K. Swamy, U. Chandra, B.S. Sherigara, "Simultaneous detection of dopamine and ascorbic acid using polyglycine modified carbon paste electrode: A cyclic voltammetric study", *J. Electroanal. Chem*. Vol. 636, pp. 80–85, 2009.